



Pregão da SES <pregao02@ses.mt.gov.br>

**QUESTIONAMENTO - PE 015/2023 - SECRETARIA ESTADUAL DE SAÚDE/MT
(UASG: 926289)**

Vitoria Giovani <Vitoria.Giovani@arjo.com>

15 de fevereiro de 2023 às 09:55

Para: "pregao02@ses.mt.gov.br" <pregao02@ses.mt.gov.br>

Cc: Keli Barris <keli.barris@arjo.com>, Guilherme Ferrari <Guilherme.Ferrari@arjo.com>

Bom dia a todos.

Prezados,

Em leitura ao edital, e interessados em participar do certame, verificamos pontos que gostaríamos de maiores esclarecimentos. À saber:

Item 12 – Cama Hospitalar*Solicitam:* RODÍZIOS DE NO MÍNIMO 5 POLEGADAS DE DIÂMETRO*Questionamos:* Gostaríamos de verificar se será considerada a medida de 12.5cm, pois aplicando a conversão simples de polegadas para cm, 5 polegadas resultariam em 12,7cm, principalmente uma vez que não existem evidências técnicas que comprovam diferença no efeito clínico por conta de somente 0.2cm de diferença.**Item 17 – Elevador***Solicitam:* RODÍZIOS DIANTEIROS 3"X2" E TRASEIROS 3"X2" COM FREIO*Questionamos:* O que a Insituição entende por rodízios de “3x2” polegadas?*Solicitam:* DRIVE MICROPROCESSADO DE 15A*Questionamos:* Serão aceitos elevadores sem drive microprocessado?*Solicitam:* LEVANTE COM ATUADOR LINEAR ELÉTRICO DE 12V*Questionamos:* Serão aceitos elevadores com atuador elétrico de 24v?*Solicitam:* BATERIA 12V 7A COM AUTONOMIA PARA 80 ACIONAMENTOS,*Questionamos:* Serão aceitas baterias de 24v e com autonomia para no mínimo 40 ciclos completos do equipamento?

Ficamos no aguardo e à disposição.

Atenciosamente,

Vitória Giovani
Sales Bid Intern

Phone: +55 (11) 5990-3139

Mobile: +55 (11) 95202-1536



Arjo Brasil

Rua Marina Ciufuli Zanfalice, 329 - Lapa

São Paulo/SP Zipcode: 05040-000

Brasil

www.arjo.com



This email and any files transmitted with it are confidential and intended solely for the use of the individual or entity to which they are addressed. If you have received this email in error please notify the system manager. This message contains confidential information and is intended only for the individual named. If you are not the named addressee you should not disseminate, distribute or copy this e-mail. Please notify the sender immediately by e-mail if you have received this e-mail by mistake and delete this e-mail from your system. If you are not the intended recipient you are notified that disclosing, copying, distributing or taking any action in reliance on the contents of this information is strictly prohibited.

This email has been scanned by the Trend Micro Email Security system
on behalf of Arjo.
